

## n-Propyl Nitrate

---

Analyte:	n-Propyl Nitrate	Method No.:	S227
Matrix:	Air	Range:	50-209 mg/cu m
OSHA Standard:	25 ppm (110 mg/cu m)	Precision ( $\overline{CV}_T$ ):	0.054
Procedure:	Adsorption on charcoal, desorption with carbon disulfide, GC	Validation Date:	12/19/75

---

### 1. Principle of the Method

- 1.1 A known volume of air is drawn through a charcoal tube to trap the organic vapors present.
- 1.2 The charcoal in the tube is transferred to a small, stoppered sample container and the analyte is desorbed with carbon disulfide.
- 1.3 An aliquot of the desorbed sample is injected into a gas chromatograph.
- 1.4 The area of the resulting peak is determined and compared with areas obtained from injection of standards.

### 2. Range and Sensitivity

- 2.1 This method was validated over the range of 50-209 mg/cu m at an atmospheric temperature and pressure of 21°C and 771 mm Hg, using a 70-liter sample. Under the conditions of sample size (70 liters) the probable useful range of this method is 11-330 mg/cu m at a detector sensitivity that gives nearly full deflection on the strip chart recorder for a 23-mg sample. This method is capable of measuring much smaller amounts if the desorption efficiency is adequate. Desorption efficiency must be determined over the range used.
- 2.2 The upper limit of the range of the method is dependent on the adsorptive capacity of the charcoal tube. This capacity varies with the concentrations of analyte and other substances in the air. The first section of the charcoal tube was found to hold 24 mg of analyte when a test atmosphere containing 209 mg/cu m of analyte in air was sampled at 0.79 liters per minute for 143 minutes; breakthrough was observed at this time, i.e., the concentration of analyte in the effluent was 5% of that in the influent.

(The charcoal tube consists of two sections of activated charcoal separated by a section of urethane foam. See Section 6.2). If a particular atmosphere is suspected of containing a large amount of contaminant, a smaller sampling volume should be taken.

### 3. Interference

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped efficiently. Preliminary experiments using toluene indicate that high humidity severely decreases the breakthrough volume.
- 3.2 When two or more compounds are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample.
- 3.3 It must be emphasized that any compound which has the same retention time as the analyte at the operating conditions described in this method is an interference. Retention time data on a single column cannot be considered as proof of chemical identity.
- 3.4 If the possibility of interference exists, separation conditions (column packing, temperature, etc.) must be changed to circumvent the problem.

### 4. Precision and Accuracy

- 4.1 The Coefficient of Variation ( $\overline{CV_T}$ ) for the total analytical and sampling method in the range of 50-209 mg/cu m was 0.054. This value corresponds to a 5.9 mg/cu m standard deviation at the OSHA standard level. Statistical information and details of the validation and experimental test procedures can be found in Reference 11.2.
- 4.2 On the average the concentrations obtained at the OSHA standard level using the overall sampling and analytical method were 0.5% higher than the "true" concentrations for a limited number of laboratory experiments. Any difference between the "found" and "true" concentrations may not represent a bias in the sampling and analytical method, but rather a random variation from the experimentally determined "true" concentration. Therefore, no recovery correction should be applied to the final result.

These data are based on validation experiments using the internal standard method.

### 5. Advantages and Disadvantages of the Method

- 5.1 The sampling device is small, portable, and involves no liquids. Interferences are minimal, and most of those which do occur can

can be eliminated by altering chromatographic conditions. The tubes are analyzed by means of a quick, instrumental method. The method can also be used for the simultaneous analysis of two or more substances suspected to be present in the same sample by simply changing gas chromatographic conditions from isothermal to a temperature-programmed mode of operation and/or by using another column.

- 5.2 One disadvantage of the method is that the amount of sample which can be taken is limited by the number of milligrams that the tube will hold before overloading. When the sample value obtained for the backup section of the charcoal tube exceeds 25% of that found on the front section, the possibility of sample loss exists.
- 5.3 Furthermore, the precision of the method is limited by the reproducibility of the pressure drop across the tubes. This drop will affect the flow rate and cause the volume to be imprecise, because the pump is usually calibrated for one tube only.

## 6. Apparatus

- 6.1 A calibrated personal sampling pump whose flow can be determined within +5% at the recommended flow rate. (Reference 11.3)
- 6.2 Charcoal tubes: glass tube with both ends flame-sealed, 7 cm long with a 6-mm O.D., and a 4-mm I.D., containing 2 sections of 20/40 mesh activated charcoal separated by a 2-mm portion of urethane foam. The activated charcoal is prepared from coconut shells and is fired at 600°C prior to packing. The adsorbing section contains 100 mg of charcoal, the backup section 50 mg. A 3-mm portion of urethane foam is placed between the outlet end of the tube and the backup section. A plug of silylated glass wool is placed in front of the adsorbing section. The pressure drop across the tube must be less than one inch of mercury at a flow rate of 1 liter per minute.
- 6.3 Gas chromatograph equipped with a flame ionization detector.
- 6.4 Column (20-ft x 1/8-in stainless steel) packed with 10% FFAP stationary phase on 80/100 mesh Chromosorb W-AW.
- 6.5 An electronic integrator or some other suitable method for measuring peak areas.
- 6.6 Two-milliliter sample containers with glass stoppers or Teflon-lined caps. If an automatic sample injector is used, the associated vials may be used.
- 6.7 Microliter syringes: 10-microliter, and other convenient sizes for making standards.
- 6.8 Pipets: 1.0 ml, graduated in 0.1-ml increments.

6.9 Volumetric flasks: 10-ml or convenient sizes for making standard solutions.

7. Reagents

7.1 Chromatographic quality carbon disulfide.

7.2 n-Propylnitrate, reagent grade.

7.3 Decane, or other suitable internal standard.

7.4 Purified nitrogen.

7.5 Prepurified hydrogen.

7.6 Filtered compressed air.

8. Procedure

8.1 Cleaning of equipment. All glassware used for the laboratory analysis should be detergent washed and thoroughly rinsed with tap water and distilled water.

8.2 Calibration of Personal Pumps. Each personal pump must be calibrated with a representative charcoal tube in the line. This will minimize errors associated with uncertainties in the sample volume collected.

8.3 Collection and Shipping of Samples

8.3.1 Immediately before sampling, break the ends of the tube to provide an opening at least one-half the internal diameter of the tube (2 mm).

8.3.2 The smaller section of charcoal is used as a backup and should be positioned nearest the sampling pump.

8.3.3 The charcoal tube should be placed in a vertical direction during sampling to minimize channeling through the charcoal.

8.3.4 Air being sampled should not be passed through any hose or tubing before entering the charcoal tube.

8.3.5 A sample size of 70 liters is recommended. Sample at a flow of 1.0 liter per minute or less. The flow rates should be known with an accuracy of at least +5%.

8.3.6 The temperature and pressure of the atmosphere being sampled should be recorded. If pressure reading is not available, record the elevation.

- 8.3.7 The charcoal tubes should be capped with the supplied plastic caps immediately after sampling. Under no circumstances should rubber caps be used.
- 8.3.8 One tube should be handled in the same manner as the sample tube (break, seal, and transport), except that no air is sampled through this tube. This tube should be labeled as a blank.
- 8.3.9 Capped charcoal tubes should be packed tightly and padded before they are shipped to minimize tube breakage during shipping.
- 8.3.10 A sample of the bulk material should be submitted to the laboratory in a glass container with a Teflon-lined cap. This sample should not be transported in the same container as the charcoal tubes.

#### 8.4 Analysis of Samples

- 8.4.1 Preparation of Samples. In preparation for analysis, each charcoal tube is scored with a file in front of the first section of charcoal and broken open. The glass wool is removed and discarded. The charcoal in the first (larger) section is transferred to a 2-ml stoppered sample container. The separating section of foam is removed and discarded; the second section is transferred to another stoppered container. These two sections are analyzed separately.
- 8.4.2 Desorption of Samples. Prior to analysis, 1.0 ml of carbon disulfide is pipetted into each sample container. (All work with carbon disulfide should be performed in a hood because of its high toxicity.) Desorption should be done for 30 minutes. Tests indicate that this is adequate if the sample is agitated occasionally during this period. If an automatic sample injector is used, the sample vials should be capped as soon as the solvent is added to minimize volatilization. For the internal standard method, desorb using 1.0 ml of carbon disulfide containing a known amount of the chosen internal standard.
- 8.4.3 GC Conditions. The typical operating conditions for the gas chromatograph are:
1. 30 ml/min (60 psig) Nitrogen carrier gas flow
  2. 35 ml/min (25 psig) Hydrogen gas flow to detector
  3. 400 ml/min (60 psig) Air flow to detector
  4. 225°C injector temperature
  5. 250°C manifold temperature (detector)
  6. 85°C column temperature

8.4.4 Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from blow back or distillation within the syringe needle, one should employ the solvent flush injection technique. The 10-microliter syringe is first flushed with solvent several times to wet the barrel and plunger. Three microliters of solvent are drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent, and the plunger is pulled back about 0.2 microliter to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 5-microliter aliquot is withdrawn, taking into consideration the volume of the needle, since the sample in the needle will be completely injected. After the needle is removed from the sample and prior to injection, the plunger is pulled back 1.2 microliters to minimize evaporation of the sample from the tip of the needle. Observe that the sample occupies 4.9-5.0 microliters in the barrel of the syringe. Duplicate injections of each sample and standard should be made. No more than a 3% difference in area is to be expected. An automatic sample injector can be used if it is shown to give reproducibility at least as good as the solvent flush technique.

8.4.5 Measurement of area. The area of the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

## 8.5 Determination of Desorption Efficiency

8.5.1 Importance of determination. The desorption efficiency of a particular compound can vary from one laboratory to another and also from one batch of charcoal to another. Thus, it is necessary to determine at least once the percentage of the specific compound that is removed in the desorption process, provided the same batch of charcoal is used.

8.5.2 Procedure for determining desorption efficiency. Activated charcoal equivalent to the amount in the first section of the sampling tube (100 mg) is measured into a 2.5 in, 4-mm I.D. glass tube, flame sealed at one end. This charcoal must be from the same batch as that used in obtaining the samples and can be obtained from unused charcoal tubes. The open end is capped with Parafilm. A known amount of the analyte is injected directly into the activated charcoal with a microliter syringe, and the tube is capped with more Parafilm. When using an automatic sample injector, the sample injector vials, capped with Teflon-faced septa, may be used in place of the glass tubes.

Six tubes at each of three concentration levels (0.5X, 1X and 2X of the standard) are prepared by adding an amount of analyte equivalent to that present in a 70-liter sample at the selected level. The tubes are allowed to stand for at least overnight to assure complete adsorption of the analyte onto the charcoal. These tubes are referred to as the samples. A parallel blank tube should be treated in the same manner except that no sample is added to it. The sample and blank tubes are desorbed and analyzed in exactly the same manner as the sampling tube described in Section 8.4.

Two or three standards are prepared by injecting the same volume of compound into 1.0 ml of carbon disulfide with the same syringe used in the preparation of the samples. These are analyzed with the samples.

If the internal standard method is used, prepare calibration standards by using 1.0 ml of carbon disulfide containing a known amount of the internal standard.

The desorption efficiency (D.E.) equals the average weight in mg recovered from the tube divided by the weight in mg added to the tube, or

$$\text{D.E.} = \frac{\text{Average Weight (mg) recovered}}{\text{Weight (mg) added}}$$

The desorption efficiency is dependent on the amount of analyte collected on the charcoal. Plot the desorption efficiency versus weight of analyte found. This curve is used in Section 10.4 to correct for adsorption losses.

## 9. Calibration and Standards

It is convenient to express concentration of standards in terms of mg per 1.0 ml carbon disulfide, because samples are desorbed in this amount of carbon disulfide. The density of the analyte is used to convert mg into microliters for easy measurement with a microliter syringe. A series of standards, varying in concentration over the range of interest, is prepared and analyzed under the same GC conditions and during the same time period as the unknown sample. Curves are established by plotting concentration in mg per 1.0 ml versus peak area.

For the internal standard method, use carbon disulfide containing a predetermined amount of the internal standard. The internal standard concentration used was approximately 70% of the concentration at 2X the standard. The analyte concentration in mg per ml is plotted versus the area ratio of the analyte to that of the internal standard. Note: Whether the external standard or internal standard method is used, standard solutions should be analyzed at the same time the sample analysis is done. This will minimize the effect of variations in FID response.

## 10. Calculations

10.1 Read the weight, in mg, corresponding to each peak area from the standard curve. No volume corrections are needed, because the standard curve is based on mg per 1.0 ml carbon disulfide and the volume of sample injected is identical to the volume of the standards injected.

10.2 Corrections for the blank must be made for each sample.

$$\text{mg} = \text{mg sample} - \text{mg blank}$$

where:

mg sample = mg found in front section of sample tube

mg blank = mg found in front section of blank tube

A similar procedure is followed for the backup sections.

10.3 Add the amounts present in the front and backup sections of the same sample tube to determine the total weight in the sample.

10.4 Read the desorption efficiency from the curve (see Section 8.5.2) for the amount found in the front section. Divide the total weight by this desorption efficiency to obtain the corrected mg/sample.

$$\text{Corrected mg/sample} = \frac{\text{Total weight}}{\text{D.E.}}$$

10.5 The concentration of the analyte in the air sampled can be expressed in mg per cu m.

$$\text{mg/cu m} = \frac{\text{Corrected mg (Section 10.4)} \times 1000 \text{ (liters/cu m)}}{\text{Air Volume Sampled (liters)}}$$

10.6 Another method of expressing concentration is ppm (corrected to standard conditions of 25°C and 760 mm Hg).

$$\text{ppm} = \text{mg/cu m} \times \frac{24.45}{\text{MW}} \times \frac{760}{\text{P}} \times \frac{(\text{T} + 273)}{298}$$

where:

P = pressure (mm Hg) of air sampled  
T = temperature (°C) of air sampled  
24.45 = molar volume (liter/mole) at 25°C and 760 mm Hg  
MW = molecular weight  
760 = standard pressure (mm Hg)  
298 = standard temperature (°K)



11. References

- 11.1 White, L.D. et al, "A Convenient Optimized Method for the Analysis of Selected Solvent Vapors in the Industrial Atmosphere," Amer. Ind. Hyg. Assoc. J., 31: 225 (1970).
- 11.2 Documentation of NIOSH Validation Tests, NIOSH Contract No. CDC-99-74-45.
- 11.3 Final Report, NIOSH Contract HSM-99-71-31, "Personal Sampler Pump for Charcoal Tubes," September 15, 1972.